

MEDI-RDT MOLD INSTRUCTION MANUAL



Disclaimer: For use in prescription pharmacy compounding only, in accordance with applicable law. MEDISCA makes no warranty or representation regarding the use of this product or fitness for any particular purpose, and any such use is entirely the responsibility of the pharmacist in the exercise of his/her professional expertise and judgment. Accordingly, in no event will MEDISCA be liable for any damages from the use or misuse of the product. Purchaser's sole remedy in the event of product defect or other damages shall be the refund of the purchase price of the product.

This manual is intended for use by qualified compounding personnel engaged in the service, inspection, maintenance and handling of the equipment.





INDEX

| SECTION A: Medi-RDT Molds | 4 - 5 |
|--|--------|
| SECTION B: Preparatory Instruction for Compounding RDTs Containing API | 5 - 6 |
| SECTION C : Filling of Molds | 7 |
| SECTION D : Baking | ε |
| SECTION E : Removal of RDTs from Mold | 9 - 10 |
| SECTION F : Dispensing RDTs | 10 |
| SECTION G : Mold Cleaning and Maintenance | 11 |
| SECTION H : Sample / Blank Formula | 1 |


CALL NOW OR VISIT OUR WEBSITE TO PLACE AN ORDER

1.800.665.6334 | WWW.MEDISCA.COM | 1.800.932.1039




MEDISCA's new RDT Molds are developed to provide quick, convenient and consistent medicated dosages. When used in conjunction with MEDISCA's Medi-RDT Base, they provide an easier and efficient way to dispense your RDTs with sufficient mechanical strength and quick disintegration in the mouth. Depending on your various compounding needs, MEDISCA offers two different customized molds.

SECTION A: MEDI-RDT MOLDS



Picture 1



Picture 2

Mold, Medi-RDT, PC, 96 Cavities, 750mg (Prod No: 7314):

- 1. Base Cavity Plate
- 2. Top Compression Plate
- 3. Lateral Containment System

Mold, Medi-RDT, PC, 180 Cavities, MultiDose 75, 150, 200mg (Prod No: 7311):

- 1. Base Cavity Plate
- 2. 60 x 200mg Cavities
- 3. 60 x 150mg Cavities
- 4. 60 x 75mg Cavities
- 5. Top Compression Plate for 200mg & 150mg Cavities
- 6. Top Compression Plate for 75mg Cavities



Other Devices / Equipment Required

- 1. Sieve, 50 Mesh, Full Height 2-5/8", Brass/Brass (*Product No: 7310*)
- 2. Sieve Receiver Pan, Brass, 8" Dia. (Product No: 9652)
- 3. Hot Hand Protector, Gripper, Red (Product No: 7315)
- **4.** Convection Oven, Mechanical, High Temperature, 1.9cu.ft. (*Product No: 7840*)
- 5. Blister Packs, Medi-RDT, w/Label & Sleeve (Product No: 7219)

SECTION B: PREPARATORY INSTRUCTIONS FOR COMPOUNDING RAPID DISSOLVE TABLETS CONTAINING API

- 1. Weigh API(s), Medi-RDT Base and other ingredients of formulations (Fig. 1).
 - **NOTE**: The API used in formulation must be stable at 100°C temperature. Refer to drug reference guide or call technical service department.
 - To account for processing error, it is suggested to measure an additional 5-9 % of the required quantities of ingredients to compound a prescribed formula.
 - The amount of Medi-RDT Base will vary with the API content. Refer to sample formula to correctly determine the required amount of Medi-RDT Base and other ingredients of the formulation
 - Preheat Convection Oven to 110°C.



(Fig. 1)



2. Triturate API(s) and other ingredients of the formulation (*Fig. 2*). Sieve through 50 mesh test sieve (*Fig. 3*) and mix geometrically all the ingredients of the formulation to obtain uniform mixture (*Fig. 4*). The Medi-RDT Base should not be crushed or triturated. It should only be sieved and mixed geometrically with the APIs and the other ingredients using a mortar and pestle (*Fig. 5*).



Fig. 2



Fig. 4



Fig. 3



Fig. 5



SECTION C: FILLING OF MOLDS

3. Fill the cavities of the mold with the powder mixture. To fill the mold cavities, pour approximately half of the powder mixture and spread over the cavities with the use of a powder scraper. Gently tap the cavity plate on the counter to allow the powder to settle and add more powder mixture as needed. Remove excess powder mixture through the lateral containment system and reserve it. It is important to completely fill the cavities (*Fig.* 6).



Fig. 6

4. Compact the powder in the cavity with the top compression plate. Remove the top plate and fill the cavities with the portion of remaining powder mixture from step-3. Remove excess powder and reserve it. Compact again with the top plate. Repeat above steps at least three times in order to completely fill the cavities (*Fig. 7* and *Fig. 8*).







This manual is intended for use by qualified compounding personnel engaged in the service, inspection, maintenance and handling of the equipment.



SECTION D: BAKING

5. Place the filled mold (only the base cavity plate) on the middle rack of a convection oven preheated at 110°C for 15 minutes. The **baking time must be very specific** (if necessary, use a timer for this procedure).




Fig. 9

6. Carefully remove the mold from the oven with the use of a hot hand protector.







Fig. 11



SECTION E: REMOVAL OF RDTs FROM MOLD

7. Immediately place a sheet of wax paper over the filled cavities (*Fig. 12*). Allow the mold to cool at room temperature between 1 - 2 minutes before removing the tablets. If waiting time exceeds 2 minutes, the removal of the tablets may become difficult.




Fig. 12

8. Invert (upside down) cavity plate on wax paper as shown in (*Fig. 13*). Release the tablets by gently tapping with the spatula handle or your hands against the bottom of the cavity plate (*Fig. 14*). Lift the cavity mold up allowing the tablets to remain on the wax paper.





Fig. 13 Fig. 14



9. Allow tablets to cool for an additional 15 - 30 minutes at controlled temperature and relative humidity, before packaging.




Fig. 15

SECTION F: DISPENSING RDTs

10. For unit dose packaging, transfer the Rapid Dissolve Tablet in Amber Circular Blister Sheets, attach the adhesive foil tamper labels and placed in to blister sleeve (*Medi-RDT Blister Packs – Product No.: 7219*).



Fig. 16



Fig. 17



SECTION G: MOLD CLEANING AND MAINTENANCE

- 1. Soak the molds in warm water-soap solution for a few minutes to facilitate cleaning.
- 2. Wash the molds using warm water, liquid soap and a soft brush or sponge.
- 3. Rinse and drain off cavity mold and top compression plate.
- 4. Use 40% alcohol for final wipe down.
- **5**. Let dry.

ATTENTION: Avoid use of metal utensils (spoon, spatulas, scrapper, brush, etc.) and cleaning aids with abrasive surface at any time due to potential risk of damaging coating of the molds.

Dishwasher can be used with all the above precautions.

SECTION H: SAMPLE / BLANK FORMULA



To view a Medi-RDT sample formula, visit: http://www.medisca.net/rdt